



# EC Design-Examination Certificate

Directive 93/42/EEC on Medical Devices, Annex II Section 4

**No.****CE 608298**

## Issued To:

**Argon Medical Devices, Inc.  
Also Trading as Becton Dickinson  
Infusion Therapy Systems Inc.  
1445 Flat Creek Road  
Athens  
Texas  
75751  
USA**

## In respect of:

**Atrieve™ Vascular Snare Kit**

BSI has performed a design examination on the above devices in accordance with the Council Directive 93/42/EEC, Annex II Section 4. The design conforms to the requirements of this directive. For marketing of these products an additional Annex II excluding Section 4 certificate is required.

For and on behalf of BSI, a Notified Body for the above Directive (Notified Body Number 0086):

Gary Fenton, Global Assurance Director

First Issued: **28/02/2014**Date: **28/02/2014**Expiry Date: **27/02/2019**

...making excellence a habit.™

Page 1 of 3

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.

This certificate was issued electronically and is bound by the conditions of the contract.

# EC Design-Examination Certificate

## Supplementary Information to CE 608298

Issued To:

**Argon Medical Devices, Inc.  
Also Trading as Becton Dickinson  
Infusion Therapy Systems Inc.  
1445 Flat Creek Road  
Athens  
Texas  
75751  
USA**

## Product: Atrieve™ Vascular Snare Kit

Product Code	Diameters (mm)	Working Diameter (mm)	Snare Length (cm)	Shaft Diameter (in)	Delivery Catheter Size (Fr)	Delivery Catheter Length (cm)	Delivery Catheter Inside Diameter (in)
381003004	4	2-4	175	0.018	3.2	150	0.030
381003008	8	4-8	175	0.018	3.2	150	0.030
382006010	10	6-10	120	0.026	6	100	0.062
382006015	15	9-15	120	0.026	6	100	0.062
382006020	20	12-20	120	0.026	6	100	0.062
382007030	30	18-30	120	0.026	7	100	0.074
382007045	45	27-45	120	0.026	7	100	0.074

First Issued: **28/02/2014**

Date: **28/02/2014**

Expiry Date: **27/02/2019**

Page 2 of 3

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.

This certificate was issued electronically and is bound by the conditions of the contract.

Information and Contact: BSI, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP. Tel: + 44 845 080 9000  
BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.  
A member of BSI Group of Companies.

# EC Design-Examination Certificate

## Supplementary Information to CE 608298

Issued To:

**Argon Medical Devices, Inc.  
Also Trading as Becton Dickinson  
Infusion Therapy Systems Inc.  
1445 Flat Creek Road  
Athens  
Texas  
75751  
USA**

## Certificate History

Date	Reference Number	Action
28 February 2014	10144761	First Issue



First Issued: **28/02/2014**

Date: **28/02/2014**

Expiry Date: **27/02/2019**

Page 3 of 3

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.

This certificate was issued electronically and is bound by the conditions of the contract.

Information and Contact: BSI, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP. Tel: + 44 845 080 9000  
BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.  
A member of BSI Group of Companies.